

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025429.D
 Acq On : 02 Dec 2021 15:32
 Operator : JC/MD
 Sample : PB141130ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141130ZHE#04

Quant Time: Dec 03 03:07:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110633	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	185736	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73651	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64074	54.945	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.900%	
35) Dibromofluoromethane	5.391	113	57348	50.322	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.640%	
50) Toluene-d8	8.653	98	213321	49.786	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.580%	
62) 4-Bromofluorobenzene	11.085	95	74247	45.804	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.600%	
Target Compounds						
20) Methylene Chloride	2.788	84	4227	3.470	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025429.D
Acq On : 02 Dec 2021 15:32
Operator : JC/MD
Sample : PB141130ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141130ZHE#04

Quant Time: Dec 03 03:07:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

