

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025430.D
 Acq On : 02 Dec 2021 15:55
 Operator : JC/MD
 Sample : PB141130ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141130ZHE#05

Quant Time: Dec 03 03:08:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	114081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	188773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75338	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	65488	54.460	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.920%	
35) Dibromofluoromethane	5.391	113	58607	50.600	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.200%	
50) Toluene-d8	8.653	98	218486	50.171	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.340%	
62) 4-Bromofluorobenzene	11.085	95	75608	45.893	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.780%	
Target Compounds						
20) Methylene Chloride	2.794	84	4083	3.251	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025430.D
Acq On : 02 Dec 2021 15:55
Operator : JC/MD
Sample : PB141130ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141130ZHE#05

Quant Time: Dec 03 03:08:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

