

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025431.D
 Acq On : 02 Dec 2021 16:19
 Operator : JC/MD
 Sample : M4917-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14

Quant Time: Dec 03 03:08:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107195	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	179224	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	61928	54.808	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.620%	
35) Dibromofluoromethane	5.391	113	55610	50.570	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.140%	
50) Toluene-d8	8.653	98	206960	50.056	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.120%	
62) 4-Bromofluorobenzene	11.079	95	69936	44.712	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.420%	
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	3.123	73	5840	1.687	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025431.D
Acq On : 02 Dec 2021 16:19
Operator : JC/MD
Sample : M4917-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14

Quant Time: Dec 03 03:08:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

