

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025432.D
 Acq On : 02 Dec 2021 16:42
 Operator : JC/MD
 Sample : M4917-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15

Quant Time: Dec 03 03:08:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175155	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69575	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61887	55.840	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 111.680%		
35) Dibromofluoromethane	5.391	113	54033	50.278	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 100.560%		
50) Toluene-d8	8.653	98	202735	50.173	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 100.340%		
62) 4-Bromofluorobenzene	11.085	95	69985	45.783	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 91.560%		
Target Compounds						Qvalue
16) Acetone	2.386	43	4513	7.328	ug/l	89
17) Carbon Disulfide	2.514	76	2216	0.808	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025432.D
Acq On : 02 Dec 2021 16:42
Operator : JC/MD
Sample : M4917-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15

Quant Time: Dec 03 03:08:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

