

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025446.D
 Acq On : 02 Dec 2021 22:59
 Operator : JC/MD
 Sample : VX1202WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1202WBL02

Quant Time: Dec 03 03:14:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	104390	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176402	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70509	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	60669	55.137	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.280%
35) Dibromofluoromethane	5.397	113	54150	50.030	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
50) Toluene-d8	8.653	98	202991	49.881	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.085	95	70635	45.881	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025446.D
Acq On : 02 Dec 2021 22:59
Operator : JC/MD
Sample : VX1202WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1202WBL02

Quant Time: Dec 03 03:14:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

