

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025449.D
 Acq On : 03 Dec 2021 00:09
 Operator : JC/MD
 Sample : M4917-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-29

Quant Time: Dec 14 04:50:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	183673	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163056	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75075	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	62852	55.10	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.20%
35) Dibromofluoromethane	5.391	113	56202	49.87	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.74%
50) Toluene-d8	8.653	98	211649	49.95	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.90%
62) 4-Bromofluorobenzene	11.085	95	72582	45.28	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.56%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.965	59	34451	117.35	ug/l	97
16) Acetone	2.380	43	3844	6.06	ug/l	95
19) Methyl tert-butyl Ether	3.117	73	46472	13.30	ug/l	100
22) Diisopropyl ether	3.763	45	3171	0.92	ug/l #	83
39) Methylcyclohexane	7.385	83	695	0.34	ug/l #	88
40) Benzene	6.044	78	410095	84.48	ug/l	99
51) 4-Methyl-2-Pentanone	8.580	43	3374	1.77	ug/l	100
52) Toluene	8.720	92	25519	8.05	ug/l	97
67) Ethyl Benzene	10.195	91	149905	26.11	ug/l	97
68) m/p-Xylenes	10.305	106	68638	29.26	ug/l	96
69) o-Xylene	10.646	106	33211	14.47	ug/l	95
73) Isopropylbenzene	10.963	105	3935	0.74	ug/l	97
78) n-propylbenzene	11.305	91	7419	1.24	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	5122	1.14	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	39603	8.91	ug/l	96
95) Naphthalene	13.780	128	16869	2.98	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025449.D
Acq On : 03 Dec 2021 00:09
Operator : JC/MD
Sample : M4917-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-29

Quant Time: Dec 14 04:50:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

