

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033183.D
 Acq On : 02 Dec 2022 14:16
 Operator : JC/MD
 Sample : N5821-09DL 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-8IDL

Quant Time: Dec 03 04:32:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198108	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65775	55.947	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.900%
35) Dibromofluoromethane	5.385	113	52799	47.275	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.540%
50) Toluene-d8	8.647	98	172360	51.349	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.700%
62) 4-Bromofluorobenzene	11.079	95	62857	42.844	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.680%
Target Compounds						
27) cis-1,2-Dichloroethene	4.495	96	1106	0.805	ug/l	85
44) Trichloroethene	7.129	130	54441	38.748	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033183.D
Acq On : 02 Dec 2022 14:16
Operator : JC/MD
Sample : N5821-09DL 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-8IDL

Quant Time: Dec 03 04:32:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

