

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033184.D
 Acq On : 02 Dec 2022 14:39
 Operator : JC/MD
 Sample : N5847-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2022-12-1

Quant Time: Dec 03 04:32:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120910	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195123	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172186	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77318	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66563	56.763	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.520%
35) Dibromofluoromethane	5.391	113	52759	47.961	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.920%
50) Toluene-d8	8.647	98	171105	51.773	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.540%
62) 4-Bromofluorobenzene	11.079	95	63081	43.655	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033184.D
Acq On : 02 Dec 2022 14:39
Operator : JC/MD
Sample : N5847-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2022-12-1

Quant Time: Dec 03 04:32:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

