

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX120222\  
 Data File : VX033188.D  
 Acq On : 02 Dec 2022 16:11  
 Operator : JC/MD  
 Sample : N5857-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 COMP-SAMPLE-TANK-10-11-12-31-33

Quant Time: Dec 03 04:33:54 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X112222W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 22 13:58:46 2022  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.556  | 168  | 113713   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.763  | 114  | 185727   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.055 | 117  | 167709   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.024 | 152  | 75797    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 61166    | 55.385   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =    | 110.780% |
| 35) Dibromofluoromethane    | 5.391  | 113   | 49330    | 47.113   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 94.220%  |
| 50) Toluene-d8              | 8.653  | 98    | 163370   | 51.941   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =    | 103.880% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 60433    | 43.938   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =    | 87.880%  |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX120222\  
Data File : VX033188.D  
Acq On : 02 Dec 2022 16:11  
Operator : JC/MD  
Sample : N5857-01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
COMP-SAMPLE-TANK-10-11-12-31-33

Quant Time: Dec 03 04:33:54 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X112222W.M  
Quant Title : SW846 8260  
QLast Update : Tue Nov 22 13:58:46 2022  
Response via : Initial Calibration

