

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033189.D
 Acq On : 02 Dec 2022 16:34
 Operator : JC/MD
 Sample : N5858-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-SAMPLE-TANK-26-27-28-30

Quant Time: Dec 03 04:34:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	114956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	187063	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61754	55.308	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.620%	
35) Dibromofluoromethane	5.385	113	49220	46.672	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.340%	
50) Toluene-d8	8.647	98	161290	50.868	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.740%	
62) 4-Bromofluorobenzene	11.079	95	60588	43.736	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033189.D
Acq On : 02 Dec 2022 16:34
Operator : JC/MD
Sample : N5858-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-SAMPLE-TANK-26-27-28-30

Quant Time: Dec 03 04:34:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

