

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033192.D
 Acq On : 02 Dec 2022 17:44
 Operator : JC/MD
 Sample : N5841-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-11D

Quant Time: Dec 03 04:34:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109195	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177612	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58779	55.428	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.860%
35) Dibromofluoromethane	5.385	113	47458	47.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	8.647	98	155843	51.806	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.620%
62) 4-Bromofluorobenzene	11.079	95	59029	44.878	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.760%
Target Compounds						
						Qvalue
44) Trichloroethene	7.135	130	1121	0.890	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033192.D
Acq On : 02 Dec 2022 17:44
Operator : JC/MD
Sample : N5841-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-11D

Quant Time: Dec 03 04:34:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

