

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033194.D
 Acq On : 02 Dec 2022 18:30
 Operator : JC/MD
 Sample : N5841-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-11S

Quant Time: Dec 03 04:35:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	112639	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180468	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74783	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60569	55.366	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.740%
35) Dibromofluoromethane	5.385	113	48503	47.673	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.340%
50) Toluene-d8	8.647	98	161057	52.730	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.460%
62) 4-Bromofluorobenzene	11.079	95	60156	45.011	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.020%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	3250	3.142	ug/l #	64
12) 1,1-Dichloroethene	2.319	96	558	0.568	ug/l #	92
21) trans-1,2-Dichloroethene	3.093	96	1344	1.195	ug/l #	82
27) cis-1,2-Dichloroethene	4.489	96	151819	118.838	ug/l	90
44) Trichloroethene	7.129	130	39883	31.161	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033194.D
Acq On : 02 Dec 2022 18:30
Operator : JC/MD
Sample : N5841-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-11S

Quant Time: Dec 03 04:35:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

