

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033195.D
 Acq On : 02 Dec 2022 18:53
 Operator : JC/MD
 Sample : N5841-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-6I

Quant Time: Dec 03 04:35:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180781	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73363	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60083	55.672	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.340%
35) Dibromofluoromethane	5.385	113	48660	47.744	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.480%
50) Toluene-d8	8.647	98	158395	51.728	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.460%
62) 4-Bromofluorobenzene	11.079	95	58576	43.753	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.500%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.495	96	1644	1.304	ug/l	91
30) Chloroform	5.093	83	1755	0.803	ug/l	97
44) Trichloroethene	7.129	130	6853	5.345	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033195.D
Acq On : 02 Dec 2022 18:53
Operator : JC/MD
Sample : N5841-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-6I

Quant Time: Dec 03 04:35:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

