

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033196.D
 Acq On : 02 Dec 2022 19:16
 Operator : JC/MD
 Sample : N5841-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-12S

Quant Time: Dec 03 04:36:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108497	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177591	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158966	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73402	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59089	56.118	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.240%
35) Dibromofluoromethane	5.385	113	47054	46.998	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.000%
50) Toluene-d8	8.647	98	155347	51.640	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.280%
62) 4-Bromofluorobenzene	11.079	95	57388	43.635	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.280%
Target Compounds						
						Qvalue
21) trans-1,2-Dichloroethene	3.099	96	491	0.453	ug/l #	81
27) cis-1,2-Dichloroethene	4.489	96	50564	41.090	ug/l	90
44) Trichloroethene	7.129	130	461	0.366	ug/l	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033196.D
Acq On : 02 Dec 2022 19:16
Operator : JC/MD
Sample : N5841-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-12S

Quant Time: Dec 03 04:36:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

