

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033197.D
 Acq On : 02 Dec 2022 19:39
 Operator : JC/MD
 Sample : N5841-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-12I

Quant Time: Dec 03 04:36:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177999	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	160474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73111	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58187	54.536	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.080%
35) Dibromofluoromethane	5.391	113	46648	46.486	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.980%
50) Toluene-d8	8.647	98	154753	51.311	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.620%
62) 4-Bromofluorobenzene	11.079	95	58168	44.127	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.260%
Target Compounds						
27) cis-1,2-Dichloroethene	4.501	96	648	0.521	ug/l	89
44) Trichloroethene	7.135	130	4481	3.550	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033197.D
Acq On : 02 Dec 2022 19:39
Operator : JC/MD
Sample : N5841-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-12I

Quant Time: Dec 03 04:36:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

