

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033201.D
 Acq On : 02 Dec 2022 21:11
 Operator : JC/MD
 Sample : N5841-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-1I

Quant Time: Dec 03 04:37:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108473	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178400	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58261	55.298	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.600%
35) Dibromofluoromethane	5.391	113	46973	46.704	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.400%
50) Toluene-d8	8.647	98	150339	49.666	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.079	95	55934	42.337	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.680%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.495	96	351	0.285	ug/l	64
44) Trichloroethene	7.129	130	2383	1.883	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033201.D
Acq On : 02 Dec 2022 21:11
Operator : JC/MD
Sample : N5841-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-1I

Quant Time: Dec 03 04:37:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

