

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033207.D
 Acq On : 02 Dec 2022 23:29
 Operator : JC/MD
 Sample : PB149377ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149377ZHE#01

Quant Time: Dec 05 01:11:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	188859	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35106	30.537	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	61.080%#
35) Dibromofluoromethane	5.391	113	37141	34.883	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	69.760%#
50) Toluene-d8	8.647	98	79792	23.776	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	47.560%#
62) 4-Bromofluorobenzene	11.079	95	61208	43.763	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.520%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12029	23.378	ug/l	99
20) Methylene Chloride	2.788	84	3370	1.899	ug/l	84
37) Ethyl Acetate	4.714	43	20209	13.606	ug/l	98
43) Isopropyl Acetate	6.342	43	92118	37.502	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033207.D
Acq On : 02 Dec 2022 23:29
Operator : JC/MD
Sample : PB149377ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149377ZHE#01

Quant Time: Dec 05 01:11:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

