

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033208.D
 Acq On : 02 Dec 2022 23:52
 Operator : JC/MD
 Sample : PB149377ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149377ZHE#02

Quant Time: Dec 05 01:11:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	191276	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35923	31.062	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	62.120%#		
35) Dibromofluoromethane	5.385	113	37944	35.187	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	70.380%#		
50) Toluene-d8	8.647	98	80104	23.547	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	47.100%#		
62) 4-Bromofluorobenzene	11.079	95	62697	44.261	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.520%		
Target Compounds						
16) Acetone	2.386	43	11912	22.975	ug/l	100
18) Methyl Acetate	2.703	43	1690	1.178	ug/l	93
20) Methylene Chloride	2.794	84	3581	2.076	ug/l	87
37) Ethyl Acetate	4.715	43	19809	13.168	ug/l	99
43) Isopropyl Acetate	6.336	43	92803	37.303	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033208.D
Acq On : 02 Dec 2022 23:52
Operator : JC/MD
Sample : PB149377ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149377ZHE#02

Quant Time: Dec 05 01:11:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

