

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033209.D
 Acq On : 03 Dec 2022 00:15
 Operator : JC/MD
 Sample : PB149377ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149377ZHE#03

Quant Time: Dec 05 01:11:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	115849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195351	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35493	30.107	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	60.220%#
35) Dibromofluoromethane	5.391	113	37518	34.066	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	68.140%#
50) Toluene-d8	8.646	98	80158	23.026	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	46.060%#
62) 4-Bromofluorobenzene	11.079	95	61997	42.854	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.700%
Target Compounds						
					Qvalue	
16) Acetone	2.391	43	11765	22.329	ug/l	96
18) Methyl Acetate	2.708	43	1688	1.157	ug/l	95
20) Methylene Chloride	2.794	84	3545	1.988	ug/l	90
37) Ethyl Acetate	4.720	43	19320	12.575	ug/l	99
43) Isopropyl Acetate	6.336	43	91195	35.892	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033209.D
Acq On : 03 Dec 2022 00:15
Operator : JC/MD
Sample : PB149377ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149377ZHE#03

Quant Time: Dec 05 01:11:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

