

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033210.D
 Acq On : 03 Dec 2022 00:38
 Operator : JC/MD
 Sample : PB149377ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149377ZHE#04

Quant Time: Dec 05 01:12:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	114652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	194967	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79324	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35104	30.086	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	60.180%#
35) Dibromofluoromethane	5.385	113	36760	33.444	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	66.880%#
50) Toluene-d8	8.647	98	78949	22.694	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	45.380%#
62) 4-Bromofluorobenzene	11.079	95	60365	41.808	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.620%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	12175	23.349	ug/l	100
18) Methyl Acetate	2.703	43	1945	1.347	ug/l	95
20) Methylene Chloride	2.794	84	3193	1.689	ug/l	92
37) Ethyl Acetate	4.715	43	18804	12.264	ug/l	99
43) Isopropyl Acetate	6.342	43	92438	36.453	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033210.D
Acq On : 03 Dec 2022 00:38
Operator : JC/MD
Sample : PB149377ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149377ZHE#04

Quant Time: Dec 05 01:12:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

