

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033211.D
 Acq On : 03 Dec 2022 01:01
 Operator : JC/MD
 Sample : PB149377ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149377ZHE#05

Quant Time: Dec 05 01:12:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117157	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	197940	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81135	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	34011	28.353	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	56.700%#
35) Dibromofluoromethane	5.385	113	36925	33.090	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	66.180%#
50) Toluene-d8	8.647	98	78510	22.182	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	44.360%#
62) 4-Bromofluorobenzene	11.079	95	60856	41.515	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.040%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	9550	17.923	ug/l	98
18) Methyl Acetate	2.703	43	1636	1.109	ug/l #	81
20) Methylene Chloride	2.788	84	4291	2.641	ug/l #	83
37) Ethyl Acetate	4.714	43	16546	10.629	ug/l	97
43) Isopropyl Acetate	6.342	43	77521	30.111	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033211.D
Acq On : 03 Dec 2022 01:01
Operator : JC/MD
Sample : PB149377ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149377ZHE#05

Quant Time: Dec 05 01:12:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

