

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033214.D
 Acq On : 03 Dec 2022 02:11
 Operator : JC/MD
 Sample : PB149377ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149377ZHE#08

Quant Time: Dec 05 01:13:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	101725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173487	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35138	34.370	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	68.740%#
35) Dibromofluoromethane	5.385	113	37942	38.793	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	77.580%
50) Toluene-d8	8.646	98	79686	26.045	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	52.080%#
62) 4-Bromofluorobenzene	11.079	95	60826	47.344	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.680%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	9763	21.102	ug/l	98
18) Methyl Acetate	2.709	43	1666	1.301	ug/l	93
20) Methylene Chloride	2.794	84	4967	3.965	ug/l #	83
37) Ethyl Acetate	4.714	43	15957	11.695	ug/l	99
43) Isopropyl Acetate	6.336	43	75985	33.675	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033214.D
Acq On : 03 Dec 2022 02:11
Operator : JC/MD
Sample : PB149377ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149377ZHE#08

Quant Time: Dec 05 01:13:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

