

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033215.D
 Acq On : 03 Dec 2022 02:34
 Operator : JC/MD
 Sample : PB149377ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149377ZHE#09

Quant Time: Dec 05 01:13:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	100873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	171947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	33561	32.982	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	65.960%#
35) Dibromofluoromethane	5.385	113	35218	36.331	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	72.660%#
50) Toluene-d8	8.647	98	76671	25.218	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	50.440%#
62) 4-Bromofluorobenzene	11.079	95	58863	46.226	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	9367	20.417	ug/l	97
18) Methyl Acetate	2.703	43	1589	1.251	ug/l	92
20) Methylene Chloride	2.788	84	4978	4.022	ug/l #	90
37) Ethyl Acetate	4.721	43	15819	11.698	ug/l	97
43) Isopropyl Acetate	6.336	43	73902	33.045	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033215.D
Acq On : 03 Dec 2022 02:34
Operator : JC/MD
Sample : PB149377ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149377ZHE#09

Quant Time: Dec 05 01:13:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

