

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033202.D
 Acq On : 02 Dec 2022 21:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 03 04:37:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.382	0.316	17.3	82	0.00
3 P	Chloromethane	0.412	0.344	16.5	84	0.00
4 C	Vinyl Chloride	0.459	0.414	9.8#	88	0.00
5 T	Bromomethane	0.340	0.318	6.5	95	0.01
6 T	Chloroethane	0.374	0.374	0.0	94	0.00
7 T	Trichlorofluoromethane	0.891	0.909	-2.0	100	0.00
8 T	Diethyl Ether	0.281	0.307	-9.3	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.481	-2.8	98	0.00
10 T	Methyl Iodide	0.536	0.478	10.8	79	0.00
11 T	Tert butyl alcohol	0.077	0.070	9.1	87	0.06
12 CM	1,1-Dichloroethene	0.436	0.446	-2.3#	98	0.00
13 T	Acrolein	0.089	0.132	-48.3#	148	0.00
14 T	Allyl chloride	0.704	0.714	-1.4	96	0.00
15 T	Acrylonitrile	0.218	0.239	-9.6	106	0.00
16 T	Acetone	0.227	0.208	8.4	87	0.01
17 T	Carbon Disulfide	1.205	0.992	17.7	82	0.00
18 T	Methyl Acetate	0.629	0.694	-10.3	108	0.00
19 T	Methyl tert-butyl Ether	1.513	1.648	-8.9	102	0.00
20 T	Methylene Chloride	0.548	0.493	10.0	96	0.00
21 T	trans-1,2-Dichloroethene	0.499	0.496	0.6	96	0.00
22 T	Diisopropyl ether	1.534	1.723	-12.3	102	0.00
23 T	Vinyl Acetate	1.201	1.317	-9.7	99	0.00
24 P	1,1-Dichloroethane	0.878	0.934	-6.4	102	0.00
25 T	2-Butanone	0.314	0.339	-8.0	103	0.00
26 T	2,2-Dichloropropane	0.706	0.694	1.7	91	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.607	-7.1	99	0.00
28 T	Bromochloromethane	0.386	0.402	-4.1	97	0.00
29 T	Tetrahydrofuran	0.199	0.222	-11.6	108	0.01
30 C	Chloroform	0.983	1.069	-8.7#	101	0.00
31 T	Cyclohexane	0.789	0.792	-0.4	94	0.00
32 T	1,1,1-Trichloroethane	0.880	0.947	-7.6	99	0.00
33 S	1,2-Dichloroethane-d4	0.539	0.464	13.9	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.282	0.254	9.9	93	0.00
36 T	1,1-Dichloropropene	0.451	0.455	-0.9	97	0.00
37 T	Ethyl Acetate	0.393	0.433	-10.2	105	0.00
38 T	Carbon Tetrachloride	0.500	0.510	-2.0	97	0.00
39 T	Methylcyclohexane	0.514	0.512	0.4	95	0.00
40 TM	Benzene	1.264	1.331	-5.3	100	0.00
41 T	Methacrylonitrile	0.221	0.248	-12.2	107	0.00
42 TM	1,2-Dichloroethane	0.509	0.548	-7.7	102	0.00
43 T	Isopropyl Acetate	0.650	0.717	-10.3	104	0.00
44 TM	Trichloroethene	0.355	0.360	-1.4	95	0.00
45 C	1,2-Dichloropropane	0.312	0.341	-9.3#	102	0.00
46 T	Dibromomethane	0.244	0.257	-5.3	101	0.00
47 T	Bromodichloromethane	0.471	0.523	-11.0	103	0.00
48 T	Methyl methacrylate	0.335	0.371	-10.7	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
 Data File : VX033202.D
 Acq On : 02 Dec 2022 21:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 03 04:37:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.008	-33.3#	119	0.04
50 S	Toluene-d8	0.915	0.759	17.0	89	0.00
51 T	4-Methyl-2-Pentanone	0.401	0.461	-15.0	108	0.00
52 CM	Toluene	0.831	0.885	-6.5#	100	0.00
53 T	t-1,3-Dichloropropene	0.462	0.524	-13.4	99	0.00
54 T	cis-1,3-Dichloropropene	0.508	0.551	-8.5	98	0.00
55 T	1,1,2-Trichloroethane	0.338	0.371	-9.8	105	0.00
56 T	Ethyl methacrylate	0.466	0.539	-15.7	103	0.00
57 T	1,3-Dichloropropane	0.546	0.600	-9.9	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.221	0.234	-5.9	100	0.00
59 T	2-Hexanone	0.296	0.340	-14.9	105	0.00
60 T	Dibromochloromethane	0.364	0.406	-11.5	100	0.00
61 T	1,2-Dibromoethane	0.357	0.395	-10.6	102	0.00
62 S	4-Bromofluorobenzene	0.370	0.346	6.5	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.354	0.337	4.8	95	0.00
65 PM	Chlorobenzene	0.997	1.041	-4.4	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.395	-8.2	102	0.00
67 C	Ethyl Benzene	1.768	1.892	-7.0#	100	0.00
68 T	m/p-Xylenes	0.692	0.734	-6.1	100	0.00
69 T	o-Xylene	0.670	0.713	-6.4	99	0.00
70 T	Styrene	1.096	1.220	-11.3	100	0.00
71 P	Bromoform	0.280	0.303	-8.2	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.573	3.732	-4.5	100	0.00
74 T	N-amyl acetate	1.215	1.325	-9.1	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.078	1.143	-6.0	104	0.00
76 T	1,2,3-Trichloropropane	0.947	1.040	-9.8	106	0.00
77 T	Bromobenzene	0.879	0.897	-2.0	99	0.00
78 T	n-propylbenzene	4.085	4.311	-5.5	99	0.00
79 T	2-Chlorotoluene	2.481	2.544	-2.5	99	0.00
80 T	1,3,5-Trimethylbenzene	3.002	3.177	-5.8	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.291	0.306	-5.2	98	0.00
82 T	4-Chlorotoluene	2.908	3.001	-3.2	98	0.00
83 T	tert-Butylbenzene	2.988	3.147	-5.3	99	0.00
84 T	1,2,4-Trimethylbenzene	2.993	3.118	-4.2	97	0.00
85 T	sec-Butylbenzene	3.574	3.765	-5.3	98	0.00
86 T	p-Isopropyltoluene	3.028	3.208	-5.9	98	0.00
87 T	1,3-Dichlorobenzene	1.648	1.668	-1.2	99	0.00
88 T	1,4-Dichlorobenzene	1.688	1.688	0.0	97	0.00
89 T	n-Butylbenzene	2.609	2.726	-4.5	96	0.00
90 T	Hexachloroethane	0.498	0.525	-5.4	100	0.00
91 T	1,2-Dichlorobenzene	1.603	1.669	-4.1	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.215	0.238	-10.7	102	0.00
93 T	1,2,4-Trichlorobenzene	0.988	1.017	-2.9	98	0.00
94 T	Hexachlorobutadiene	0.436	0.406	6.9	94	0.00
95 T	Naphthalene	3.197	3.464	-8.4	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033202.D
Acq On : 02 Dec 2022 21:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 03 04:37:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.991	1.012	-2.1	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6