



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120222\
Data File : VX033174.D
Acq On : 02 Dec 2022 10:33
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Title : SW846 8260
Last Update : Tue Nov 22 13:58:46 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	7163	PASS
75	95	30	60	55.7	19787	PASS
95	95	100	100	100.0	35512	PASS
96	95	5	9	6.7	2371	PASS
173	174	0.00	2	0.9	246	PASS
174	95	50	100	76.6	27219	PASS
175	174	5	9	8.0	2166	PASS
176	174	95	101	95.9	26098	PASS
177	176	5	9	6.2	1620	PASS