

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044053.D
 Acq On : 02 Dec 2024 10:14
 Operator : JC/MD
 Sample : VX1202MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1202MBL01

Quant Time: Dec 03 00:13:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	133810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	257802	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	232828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	103784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	114834	50.035	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.080%		
35) Dibromofluoromethane	5.379	113	87718	47.618	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 95.240%		
50) Toluene-d8	8.641	98	318342	50.133	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.260%		
62) 4-Bromofluorobenzene	11.079	95	112349	51.988	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 103.980%		

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044053.D
Acq On : 02 Dec 2024 10:14
Operator : JC/MD
Sample : VX1202MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1202MBL01

Quant Time: Dec 03 00:13:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

