

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044061.D
 Acq On : 02 Dec 2024 13:21
 Operator : JC/MD
 Sample : P5021-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Dec 03 00:18:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	120070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	229971	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	208817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	91525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	106938	51.927	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.860%			
35) Dibromofluoromethane	5.379	113	77289	47.034	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.060%			
50) Toluene-d8	8.647	98	288636	50.955	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.920%			
62) 4-Bromofluorobenzene	11.079	95	99490	51.609	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.220%			
Target Compounds						
16) Acetone	2.386	43	1592	1.683	ug/l #	87
29) Tetrahydrofuran	5.032	42	1450	1.938	ug/l #	82
30) Chloroform	5.099	83	5656	1.886	ug/l	85
44) Trichloroethene	7.123	130	499	0.276	ug/l	58
95) Naphthalene	13.780	128	2354	0.360	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044061.D
Acq On : 02 Dec 2024 13:21
Operator : JC/MD
Sample : P5021-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Dec 03 00:18:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

