

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044063.D
 Acq On : 02 Dec 2024 14:07
 Operator : JC/MD
 Sample : PB165251TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB165251TB

Quant Time: Dec 03 00:20:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	118490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	233518	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	89747	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	105828	52.073	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.140%			
35) Dibromofluoromethane	5.385	113	77332	46.345	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 92.700%			
50) Toluene-d8	8.647	98	283973	49.371	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.740%			
62) 4-Bromofluorobenzene	11.079	95	98282	50.208	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.420%			
Target Compounds						
16) Acetone	2.386	43	37961	40.669	ug/l	95
18) Methyl Acetate	2.703	43	2231	1.026	ug/l #	89
20) Methylene Chloride	2.788	84	2961	1.869	ug/l #	88
43) Isopropyl Acetate	6.336	43	115786	28.188	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044063.D
Acq On : 02 Dec 2024 14:07
Operator : JC/MD
Sample : PB165251TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB165251TB

Quant Time: Dec 03 00:20:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

