

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044064.D
 Acq On : 02 Dec 2024 14:30
 Operator : JC/MD
 Sample : P4995-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001

Quant Time: Dec 03 00:20:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	127863	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	249639	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	212809	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	89482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	113604	51.801	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.600%
35) Dibromofluoromethane	5.385	113	83644	46.891	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	8.647	98	306369	49.825	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.079	95	100715	48.128	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.260%
Target Compounds						
18) Methyl Acetate	2.709	43	3681	1.569	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044064.D
Acq On : 02 Dec 2024 14:30
Operator : JC/MD
Sample : P4995-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001

Quant Time: Dec 03 00:20:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

