

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044065.D
 Acq On : 02 Dec 2024 14:52
 Operator : JC/MD
 Sample : P5001-05DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C2-2DL

Quant Time: Dec 03 00:20:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	122610	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	241771	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	210916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	99346	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	109043	51.852	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.700%
35) Dibromofluoromethane	5.379	113	78946	45.698	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.400%
50) Toluene-d8	8.647	98	293792	49.334	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.660%
62) 4-Bromofluorobenzene	11.079	95	103588	51.112	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.220%

Target Compounds					Qvalue
27) cis-1,2-Dichloroethene	4.483	96	17444	9.675	ug/l 91
40) Benzene	6.037	78	27964	4.170	ug/l 99
44) Trichloroethene	7.129	130	2205	1.162	ug/l 58
52) Toluene	8.714	92	31614	7.877	ug/l 99
67) Ethyl Benzene	10.189	91	168318	21.443	ug/l 100
68) m/p-Xylenes	10.299	106	26583	9.081	ug/l 96
69) o-Xylene	10.640	106	21201	7.418	ug/l 92
80) 1,3,5-Trimethylbenzene	11.451	105	7643	1.217	ug/l 97
84) 1,2,4-Trimethylbenzene	11.750	105	24518	3.920	ug/l 95
95) Naphthalene	13.774	128	2178441	306.560	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044065.D
Acq On : 02 Dec 2024 14:52
Operator : JC/MD
Sample : P5001-05DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C2-2DL

Quant Time: Dec 03 00:20:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

