

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044066.D
 Acq On : 02 Dec 2024 15:15
 Operator : JC/MD
 Sample : P5001-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C2-3DL

Quant Time: Dec 03 00:21:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	120449	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	231562	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	95432	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	105705	51.166	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.340%			
35) Dibromofluoromethane	5.379	113	75420	45.581	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.160%			
50) Toluene-d8	8.647	98	287296	50.370	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.740%			
62) 4-Bromofluorobenzene	11.079	95	102472	52.790	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.580%			
Target Compounds						
27) cis-1,2-Dichloroethene	4.483	96	27228	15.372	ug/l	89
40) Benzene	6.037	78	64882	10.102	ug/l	98
44) Trichloroethene	7.129	130	3102	1.706	ug/l	87
52) Toluene	8.714	92	52917	13.766	ug/l	98
67) Ethyl Benzene	10.189	91	308936	40.625	ug/l	99
68) m/p-Xylenes	10.299	106	43816	15.451	ug/l	97
69) o-Xylene	10.640	106	34942	12.620	ug/l	91
73) Isopropylbenzene	10.963	105	11928	1.626	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	11810	1.957	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	39796	6.623	ug/l	96
95) Naphthalene	13.774	128	3349221	490.647	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044066.D
Acq On : 02 Dec 2024 15:15
Operator : JC/MD
Sample : P5001-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C2-3DL

Quant Time: Dec 03 00:21:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

