

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044067.D
 Acq On : 02 Dec 2024 15:38
 Operator : JC/MD
 Sample : P5001-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C3-2DL

Quant Time: Dec 03 00:22:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	127859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	242833	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	215764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	100587	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	112108	51.121	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.240%	
35) Dibromofluoromethane	5.379	113	82402	47.489	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.980%	
50) Toluene-d8	8.647	98	306098	51.176	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.360%	
62) 4-Bromofluorobenzene	11.079	95	106724	52.429	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	104.860%	

Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.483	96	10955	5.826	ug/l	91
40) Benzene	6.031	78	55468	8.236	ug/l	99
52) Toluene	8.714	92	92198	22.871	ug/l	99
67) Ethyl Benzene	10.189	91	307165	38.253	ug/l	99
68) m/p-Xylenes	10.299	106	55282	18.461	ug/l	100
69) o-Xylene	10.640	106	38240	13.079	ug/l	97
73) Isopropylbenzene	10.963	105	11337	1.466	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	11840	1.861	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	40565	6.405	ug/l	99
95) Naphthalene	13.774	128	3739861	519.796	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044067.D
Acq On : 02 Dec 2024 15:38
Operator : JC/MD
Sample : P5001-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C3-2DL

Quant Time: Dec 03 00:22:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

