

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044068.D
 Acq On : 02 Dec 2024 16:01
 Operator : JC/MD
 Sample : P5001-10DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C3-3DL

Quant Time: Dec 03 00:22:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	126077	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	246510	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	214521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	101769	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	110759	51.220	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.440%		
35) Dibromofluoromethane	5.379	113	79600	45.190	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 90.380%		
50) Toluene-d8	8.647	98	303198	49.935	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.880%		
62) 4-Bromofluorobenzene	11.079	95	105242	50.930	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 101.860%		
Target Compounds						
					Qvalue	
40) Benzene	6.038	78	30383	4.444	ug/l	96
52) Toluene	8.714	92	24586	6.008	ug/l	99
67) Ethyl Benzene	10.189	91	199557	24.996	ug/l	98
68) m/p-Xylenes	10.299	106	29052	9.758	ug/l	97
69) o-Xylene	10.640	106	23794	8.185	ug/l	98
73) Isopropylbenzene	10.963	105	8050	1.029	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	7474	1.161	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	24457	3.817	ug/l	97
95) Naphthalene	13.774	128	2054202	282.194	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044068.D
Acq On : 02 Dec 2024 16:01
Operator : JC/MD
Sample : P5001-10DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C3-3DL

Quant Time: Dec 03 00:22:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

