

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044070.D
 Acq On : 02 Dec 2024 16:47
 Operator : JC/MD
 Sample : P5001-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C1-2RE

Quant Time: Dec 03 00:23:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	119853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	235670	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	212261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	96574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	106814	51.960	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.920%			
35) Dibromofluoromethane	5.379	113	56815	33.738	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 67.480%#			
50) Toluene-d8	8.646	98	292627	50.411	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.820%			
62) 4-Bromofluorobenzene	11.079	95	102008	51.635	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.280%			
Target Compounds						
16) Acetone	2.385	43	9603	10.171	ug/l #	89
27) cis-1,2-Dichloroethene	4.489	96	4916	2.789	ug/l	89
40) Benzene	6.031	78	158872	24.305	ug/l	98
52) Toluene	8.713	92	225458	57.628	ug/l	100
67) Ethyl Benzene	10.189	91	1061982	134.436	ug/l	99
68) m/p-Xylenes	10.299	106	189863	64.450	ug/l	98
69) o-Xylene	10.640	106	130517	45.377	ug/l	95
73) Isopropylbenzene	10.963	105	46835	6.309	ug/l	98
78) n-propylbenzene	11.304	91	10194	1.230	ug/l #	53
80) 1,3,5-Trimethylbenzene	11.451	105	42016	6.879	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	155368	25.552	ug/l	98
95) Naphthalene	13.780	128	12096197	1751.089	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044070.D
Acq On : 02 Dec 2024 16:47
Operator : JC/MD
Sample : P5001-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C1-2RE

Quant Time: Dec 03 00:23:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

