

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044072.D
 Acq On : 02 Dec 2024 17:32
 Operator : JC/MD
 Sample : P5001-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C2-1RE

Quant Time: Dec 03 00:25:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	119282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	230844	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	204538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	97308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	103768	50.720	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.440%			
35) Dibromofluoromethane	5.379	113	47734	28.939	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 57.880%#			
50) Toluene-d8	8.647	98	286656	50.415	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.820%			
62) 4-Bromofluorobenzene	11.079	95	101529	52.467	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.940%			
Target Compounds						
						Qvalue
16) Acetone	2.380	43	26351	28.043	ug/l	96
27) cis-1,2-Dichloroethene	4.477	96	92808	52.909	ug/l	91
40) Benzene	6.038	78	47730	7.455	ug/l	99
44) Trichloroethene	7.123	130	35225	19.434	ug/l	91
52) Toluene	8.714	92	59811	15.608	ug/l	99
67) Ethyl Benzene	10.189	91	328471	43.151	ug/l	99
68) m/p-Xylenes	10.299	106	54288	19.124	ug/l	96
69) o-Xylene	10.640	106	56249	20.295	ug/l	98
73) Isopropylbenzene	10.964	105	21545	2.880	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	27542	4.476	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	100490	16.402	ug/l	97
95) Naphthalene	13.780	128	8401769	1207.096	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044072.D
Acq On : 02 Dec 2024 17:32
Operator : JC/MD
Sample : P5001-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C2-1RE

Quant Time: Dec 03 00:25:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

