

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044073.D
 Acq On : 02 Dec 2024 17:55
 Operator : JC/MD
 Sample : P5001-08RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C3-1RE

Quant Time: Dec 03 00:26:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	113564	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	222024	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	199117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	92954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99600	51.134	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.260%		
35) Dibromofluoromethane	5.379	113	48239	30.406	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 60.820%#		
50) Toluene-d8	8.647	98	273748	50.057	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.120%		
62) 4-Bromofluorobenzene	11.079	95	98038	52.676	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 105.360%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	17025	19.031	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	11953	7.157	ug/l	91
40) Benzene	6.038	78	16776	2.724	ug/l	92
44) Trichloroethene	7.123	130	3536	2.028	ug/l	94
52) Toluene	8.714	92	31239	8.476	ug/l	96
67) Ethyl Benzene	10.189	91	245622	33.146	ug/l	97
68) m/p-Xylenes	10.299	106	41834	15.138	ug/l	99
69) o-Xylene	10.640	106	35799	13.268	ug/l	97
73) Isopropylbenzene	10.964	105	15964	2.234	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	16130	2.744	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	62289	10.643	ug/l	96
95) Naphthalene	13.774	128	5811084	873.995	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044073.D
Acq On : 02 Dec 2024 17:55
Operator : JC/MD
Sample : P5001-08RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C3-1RE

Quant Time: Dec 03 00:26:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

