

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044074.D
 Acq On : 02 Dec 2024 18:18
 Operator : JC/MD
 Sample : P5001-11RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C4-1RE

Quant Time: Dec 03 00:26:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	118857	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	232004	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	88047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	106693	52.336	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.680%		
35) Dibromofluoromethane	5.385	113	57154	34.476	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 68.960%#		
50) Toluene-d8	8.647	98	282050	49.356	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 98.720%		
62) 4-Bromofluorobenzene	11.079	95	99442	51.132	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.260%		
Target Compounds						
						Qvalue
16) Acetone	2.380	43	15870	16.949	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	7451	4.263	ug/l	87
67) Ethyl Benzene	10.195	91	16960	2.262	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	5778	1.042	ug/l	97
95) Naphthalene	13.774	128	657879	104.460	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044074.D
Acq On : 02 Dec 2024 18:18
Operator : JC/MD
Sample : P5001-11RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C4-1RE

Quant Time: Dec 03 00:26:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

