

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044075.D
 Acq On : 02 Dec 2024 18:41
 Operator : JC/MD
 Sample : PB165299TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB165299TB

Quant Time: Dec 03 00:27:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	113073	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	228545	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	195726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	83157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	104660	53.965	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.940%			
35) Dibromofluoromethane	5.385	113	74338	45.520	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.040%			
50) Toluene-d8	8.647	98	278277	49.433	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.860%			
62) 4-Bromofluorobenzene	11.079	95	95862	50.037	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.080%			
Target Compounds						
95) Naphthalene	13.774	128	33614	5.651	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044075.D
Acq On : 02 Dec 2024 18:41
Operator : JC/MD
Sample : PB165299TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB165299TB

Quant Time: Dec 03 00:27:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

