

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044076.D
 Acq On : 02 Dec 2024 19:04
 Operator : JC/MD
 Sample : P5006-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C4-2

Quant Time: Dec 03 00:28:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	114189	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	228244	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	210148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	88412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	105289	53.759	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.520%			
35) Dibromofluoromethane	5.379	113	58066	35.603	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 71.200%#			
50) Toluene-d8	8.647	98	288882	51.385	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.760%			
62) 4-Bromofluorobenzene	11.079	95	103869	54.288	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 108.580%			
Target Compounds						
16) Acetone	2.380	43	28300	31.461	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	13424	7.994	ug/l	92
40) Benzene	6.037	78	12620	1.994	ug/l	99
67) Ethyl Benzene	10.195	91	52600	6.726	ug/l	99
69) o-Xylene	10.640	106	5293	1.859	ug/l	92
84) 1,2,4-Trimethylbenzene	11.750	105	17660	3.172	ug/l	99
95) Naphthalene	13.774	128	1420906	224.685	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044076.D
Acq On : 02 Dec 2024 19:04
Operator : JC/MD
Sample : P5006-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C4-2

Quant Time: Dec 03 00:28:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

