

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
 Data File : VX044077.D
 Acq On : 02 Dec 2024 19:27
 Operator : JC/MD
 Sample : P5006-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C4-3

Quant Time: Dec 03 00:28:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	116858	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	228656	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203255	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104277	52.026	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.060%		
35) Dibromofluoromethane	5.379	113	60617	37.100	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 74.200%#		
50) Toluene-d8	8.647	98	281785	50.032	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.060%		
62) 4-Bromofluorobenzene	11.079	95	99767	52.050	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 104.100%		
Target Compounds						
						Qvalue
16) Acetone	2.380	43	18162	19.729	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	2586	1.505	ug/l	89
67) Ethyl Benzene	10.195	91	24208	3.200	ug/l	92
84) 1,2,4-Trimethylbenzene	11.750	105	12720	2.244	ug/l	99
95) Naphthalene	13.774	128	1191529	185.063	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120224\
Data File : VX044077.D
Acq On : 02 Dec 2024 19:27
Operator : JC/MD
Sample : P5006-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C4-3

Quant Time: Dec 03 00:28:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

