

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX120318\
 Data File : VX006247.D
 Acq On : 03 Dec 2018 12:31
 Operator : JC/MD
 Sample : J6139-01
 Misc : 6.24g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0483

Quant Time: Dec 04 05:47:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	791638	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1170339	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1081609	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	593524	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	379178	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
35) Dibromofluoromethane	5.51	113	371242	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
50) Toluene-d8	8.71	98	1363235	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	11.14	95	537453	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	25252	2.966	ug/l	87
44) Trichloroethene	7.21	130	38752	4.334	ug/l	93
64) Tetrachloroethene	9.34	164	8901	1.093	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX120318\
Data File : VX006247.D
Acq On : 03 Dec 2018 12:31
Operator : JC/MD
Sample : J6139-01
Misc : 6.24g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
S49-0483

Quant Time: Dec 04 05:47:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

