

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX120318\
 Data File : VX006262.D
 Acq On : 03 Dec 2018 19:11
 Operator : JC/MD
 Sample : J5770-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-03-112918-A

Quant Time: Dec 04 06:14:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	815983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1239614	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1103906	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	518922	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	405284	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
35) Dibromofluoromethane	5.51	113	395515	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
50) Toluene-d8	8.71	98	1440483	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
62) 4-Bromofluorobenzene	11.13	95	521581	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
Target Compounds						
16) Acetone	2.45	43	40241	10.294	ug/l	96
20) Methylene Chloride	2.86	84	444568	53.436	ug/l	98
43) Isopropyl Acetate	6.46	43	48121	2.469	ug/l	98
64) Tetrachloroethene	9.34	164	9001	1.083	ug/l	91
95) Naphthalene	13.83	128	103624	2.600	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX120318\
Data File : VX006262.D
Acq On : 03 Dec 2018 19:11
Operator : JC/MD
Sample : J5770-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-03-112918-A

Quant Time: Dec 04 06:14:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

