

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120319\
 Data File : VX013720.D
 Acq On : 03 Dec 2019 13:23
 Operator : JC/SP
 Sample : PB125179TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125179TB

Quant Time: Dec 03 15:47:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	701778	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1067749	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	964043	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	439542	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	385626	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
35) Dibromofluoromethane	5.49	113	322010	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
50) Toluene-d8	8.71	98	1274831	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
62) 4-Bromofluorobenzene	11.13	95	440209	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
Target Compounds						
16) Acetone	2.43	43	18287	4.547	ug/l	99
20) Methylene Chloride	2.85	84	15939	2.063	ug/l	97
43) Isopropyl Acetate	6.43	43	169907	9.491	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	16138	0.639	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120319\
Data File : VX013720.D
Acq On : 03 Dec 2019 13:23
Operator : JC/SP
Sample : PB125179TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125179TB

Quant Time: Dec 03 15:47:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

