

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120319\
 Data File : VX013723.D
 Acq On : 03 Dec 2019 14:38
 Operator : JC/SP
 Sample : K5675-09
 Misc : 4.28g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-05-112619

Quant Time: Dec 03 15:07:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	639002	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	977443	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	871976	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	434712	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	355201	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
35) Dibromofluoromethane	5.49	113	282495	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
50) Toluene-d8	8.71	98	1160956	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	11.14	95	407513	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
Target Compounds						
16) Acetone	2.48	43	12105	3.306	ug/l #	82
25) 2-Butanone	4.70	43	16617	3.053	ug/l	91
84) 1,2,4-Trimethylbenzene	11.81	105	10577	0.417	ug/l	100
95) Naphthalene	13.83	128	473290	16.072	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120319\
Data File : VX013723.D
Acq On : 03 Dec 2019 14:38
Operator : JC/SP
Sample : K5675-09
Misc : 4.28g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-05-112619

Quant Time: Dec 03 15:07:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

