

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120319\
 Data File : VX013726.D
 Acq On : 03 Dec 2019 15:48
 Operator : JC/SP
 Sample : K6027-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14

Manual Integrations
 APPROVED

apatel
 12/4/2019 1:52:18 PM

Quant Time: Dec 04 06:23:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	687623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1062251	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	967941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	478768	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	372099	46.62	ug/l	0.00
Spiked Amount			Recovery	=	93.24%	
35) Dibromofluoromethane	5.49	113	316470	48.19	ug/l	0.00
Spiked Amount			Recovery	=	96.38%	
50) Toluene-d8	8.71	98	1262220	49.33	ug/l	0.00
Spiked Amount			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	11.14	95	465559	50.46	ug/l	0.00
Spiked Amount			Recovery	=	100.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Cyclohexane	5.61	56	67759	5.747	ug/l	# 47
39) Methylcyclohexane	7.46	83	149915	12.578	ug/l	93
52) Toluene	8.78	92	12045	0.660	ug/l	96
67) Ethyl Benzene	10.25	91	739859	21.424	ug/l	100
68) m/p-Xylenes	10.35	106	40481	3.062	ug/l	100
69) o-Xylene	10.70	106	105672	8.195	ug/l	98
73) Isopropylbenzene	11.01	105	2956241	88.026	ug/l	100
78) n-propylbenzene	11.35	91	3073477	81.982	ug/l	100
80) 1,3,5-Trimethylbenzene	11.50	105	966330	35.107	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	13299762	476.095	ug/l	76
85) sec-Butylbenzene	11.94	105	134726m	4.217	ug/l	
86) p-Isopropyltoluene	12.06	119	470062m	15.814	ug/l	
95) Naphthalene	13.82	128	23851486m	735.438	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120319\
Data File : VX013726.D
Acq On : 03 Dec 2019 15:48
Operator : JC/SP
Sample : K6027-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14

Quant Time: Dec 04 06:23:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

Manual Integrations
APPROVED

apatel
12/4/2019 1:52:18 PM

