

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120319\
 Data File : VX013731.D
 Acq On : 03 Dec 2019 17:44
 Operator : JC/SP
 Sample : K5984-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S39-0196

Quant Time: Dec 04 12:05:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	691329	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1053003	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	935736	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	424793	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	377848	47.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.18%	
35) Dibromofluoromethane	5.49	113	312473	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
50) Toluene-d8	8.71	98	1246782	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
62) 4-Bromofluorobenzene	11.14	95	424650	46.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.86%	
Target Compounds						
					Qvalue	
16) Acetone	2.43	43	8805	2.222	ug/l	94
21) trans-1,2-Dichloroethene	3.16	96	22388	3.124	ug/l	91
27) cis-1,2-Dichloroethene	4.59	96	266622	32.393	ug/l	89
44) Trichloroethene	7.21	130	46708	5.846	ug/l	92
95) Naphthalene	13.83	128	12794	0.445	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120319\
Data File : VX013731.D
Acq On : 03 Dec 2019 17:44
Operator : JC/SP
Sample : K5984-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S39-0196

Quant Time: Dec 04 12:05:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

