

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025472.D
 Acq On : 03 Dec 2021 11:05
 Operator : JC/MD
 Sample : VX1203WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1203WBL01

Quant Time: Dec 04 06:46:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195683	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176025	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80507	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67864	56.606	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.220%
35) Dibromofluoromethane	5.391	113	60221	50.157	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.320%
50) Toluene-d8	8.653	98	225103	49.865	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.720%
62) 4-Bromofluorobenzene	11.079	95	78991	46.253	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025472.D
Acq On : 03 Dec 2021 11:05
Operator : JC/MD
Sample : VX1203WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1203WBL01

Quant Time: Dec 04 06:46:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

