

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025475.D
 Acq On : 03 Dec 2021 12:25
 Operator : JC/MD
 Sample : M4917-19DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-35DL

Quant Time: Dec 04 06:48:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	186532	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64632	55.658	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.320%
35) Dibromofluoromethane	5.391	113	57705	50.419	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.840%
50) Toluene-d8	8.653	98	214711	49.896	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.800%
62) 4-Bromofluorobenzene	11.085	95	74671	45.869	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.740%
Target Compounds						
11) Tert butyl alcohol	2.959	59	17791	54.960	ug/l	Qvalue 98
16) Acetone	2.379	43	1192	1.847	ug/l	98
19) Methyl tert-butyl Ether	3.123	73	15993	4.495	ug/l	98
22) Diisopropyl ether	3.769	45	3061	0.870	ug/l #	81
40) Benzene	6.043	78	211465	42.894	ug/l	96
52) Toluene	8.720	92	1676	0.521	ug/l	99
67) Ethyl Benzene	10.195	91	21431	3.651	ug/l	100
68) m/p-Xylenes	10.311	106	977	0.407	ug/l	85
73) Isopropylbenzene	10.963	105	4199	0.784	ug/l	99
78) n-propylbenzene	11.305	91	3535	0.588	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025475.D
Acq On : 03 Dec 2021 12:25
Operator : JC/MD
Sample : M4917-19DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-35DL

Quant Time: Dec 04 06:48:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

