

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025482.D
 Acq On : 03 Dec 2021 15:09
 Operator : JC/MD
 Sample : M4918-04DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-39DL

Quant Time: Dec 04 06:50:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		5.556	168	102612	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		6.763	114	175920	50.000	ug/l	0.00
63) Chlorobenzene-d5		10.055	117	158585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.024	152	74077	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		5.958	65	61253	56.632	ug/l	0.00
Spiked Amount 50.000		Range	78 - 117	Recovery	=	113.260%	
35) Dibromofluoromethane		5.391	113	53932	49.965	ug/l	0.00
Spiked Amount 50.000		Range	75 - 124	Recovery	=	99.940%	
50) Toluene-d8		8.653	98	202079	49.793	ug/l	0.00
Spiked Amount 50.000		Range	92 - 112	Recovery	=	99.580%	
62) 4-Bromofluorobenzene		11.079	95	72668	47.331	ug/l	0.00
Spiked Amount 50.000		Range	83 - 123	Recovery	=	94.660%	
Target Compounds							
							Qvalue
19) Methyl tert-butyl Ether		3.123	73	1739	0.525	ug/l	# 73
39) Methylcyclohexane		7.385	83	910	0.462	ug/l	89
40) Benzene		6.044	78	52093	11.204	ug/l	100
52) Toluene		8.720	92	89231	29.396	ug/l	97
67) Ethyl Benzene		10.195	91	127939	22.909	ug/l	99
68) m/p-Xylenes		10.305	106	143831	63.054	ug/l	93
69) o-Xylene		10.646	106	85764	38.414	ug/l	92
73) Isopropylbenzene		10.963	105	5271	0.998	ug/l	99
78) n-propylbenzene		11.305	91	15757	2.659	ug/l	97
80) 1,3,5-Trimethylbenzene		11.451	105	20107	4.531	ug/l	96
84) 1,2,4-Trimethylbenzene		11.756	105	88272	20.126	ug/l	95
89) n-Butylbenzene		12.317	91	1332	0.335	ug/l	# 13
95) Naphthalene		13.780	128	37720	7.226	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025482.D
Acq On : 03 Dec 2021 15:09
Operator : JC/MD
Sample : M4918-04DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-39DL

Quant Time: Dec 04 06:50:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

