

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025487.D
 Acq On : 03 Dec 2021 17:06
 Operator : JC/MD
 Sample : M4918-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-42

Quant Time: Dec 04 06:52:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	188084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65641	56.392	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.780%	
35) Dibromofluoromethane	5.391	113	58734	50.895	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.800%	
50) Toluene-d8	8.653	98	217464	50.119	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.240%	
62) 4-Bromofluorobenzene	11.079	95	74979	45.678	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.360%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	6194	9.576	ug/l	100
25) 2-Butanone	4.568	43	2469	2.640	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025487.D
Acq On : 03 Dec 2021 17:06
Operator : JC/MD
Sample : M4918-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-42

Quant Time: Dec 04 06:52:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

